

CALIBRATION LABORATORIES

NVLAP LAB CODE 200898-0

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Rudolph Research Analytical Corporation 55 Newburgh Road Hackettstown, NJ 07840 Heidi Spanier Phone: 973-584-1558 Fax 973-584-5440 E-mail: hspanier@rudolphresearch.com URL: www.rudolphresearch.com	Fields of Calibration Dimensional Mechanical Thermodynamic Optical Radiation
--	---

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <small>Notes 1,2</small>			
Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <small>Note 3, 8</small>	Remarks
DIMENSIONAL			
ANGLE OF OPTICAL ROTATION (20/D19)			
Angle of Optical Rotation			
Quartz Control Plates <small>note 7</small>	-41° to < -30°	0.0030°	Indirect Comparison to Standard Quartz Control Plate
(546.2271 nm)	-30° to 41°	0.0020°	
Polarimeters			Direct Comparison to Quartz Control Plate
(546 nm, 589 nm, 633 nm)	(-41 to -29)°	(-0.00013α + 0.00176)°	
Field Service Available <small>Note 4</small>	(> -29 to -24)°	(-0.00012α + 0.00163)°	
	(> -24 to -15)°	(-0.00011α + 0.00183)°	
	(> -15 to -4)°	(-0.00007α + 0.00242)°	
	(> -4 to 4)°	0.0026°	
	(> 4 to 16)°	(0.00007α + 0.00242)°	
	(>16 to 24)°	(0.00011α + 0.00183)°	
	(>24 to 29)°	(0.00012α + 0.00163)°	
	(>29 to 32)°	(0.00013α + 0.00129)°	
	(>32 to 39)°	(0.00013α + 0.00135)°	
	(>39 to 41)°	(0.00014α + 0.00109)°	
Specific Rotation of Liquids	-30° to 41°	0.0069° Uncertainty is dependent on fluid identity	Indirect Comparison to Quartz Control Plates and Liquid Density Standards

2024-03-06 through 2025-03-31

Effective dates



For the National Voluntary Laboratory Accreditation Program

CALIBRATION LABORATORIES

NVLAP LAB CODE 200898-0

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <small>Notes 1,2</small>			
Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <small>Note 3, 8</small>	Remarks
MECHANICAL			
VOLUME AND DENSITY (20/M12)			
Density			
Density Meters Field Service Available <small>Note 4</small>	(0.65 to 1.8) g/cm ³	0.000011 g/cm ³	
Liquids	(0.65 to 1.8) g/cm ³	0.000019 g/cm ³ Uncertainty is dependent on fluid identity	Indirect comparison to liquid reference standards
THERMODYNAMIC			
LABORATORY THEROMETERS, DIGITAL AND ANALOG (20/T03)			
Temperature - Generate	(20 to 30) °C	0.007 °C	Direct Comparison to Fluke 5642 Probe with 1504 Meter
OPTICAL RADIATION			
SPECTROPHOTOMETRIC (20/O04)			
Refractive Index Sucrose (Brix) Solutions	(0 to 67.5) %(m/m)	0.0064 %(m/m) Uncertainty is Dependent on Mass Fraction	Gravimetric Formulation
	Refractive Index of: 1.332986 to 1.459413 Based on %(m/m)	0.000012 Uncertainty is Dependent on Mass Fraction	Refractive Index Calculated From Mass Fraction According to ICUMSA SPS-3 (2000)
END			

2024-03-06 through 2025-03-31

Effective dates



For the National Voluntary Laboratory Accreditation Program

CALIBRATION LABORATORIES

NVLAP LAB CODE 200898-0

Notes

Note 1: A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

Note 2: Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

Note 3: The uncertainty associated with a measurement in a CMC is an expanded uncertainty with a level of confidence of approximately 95 %, typically using a coverage factor of $k = 2$. However, laboratories may report a coverage factor different than $k = 2$ to achieve the 95 % level of confidence. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

Note 3a: The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

Note 3b: As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

Note 3c: As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.5 of NIST Handbook 150, Procedures and General Requirements.

Note 4: Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

Note 5: Values listed with percent (%) are percent of reading or generated value unless otherwise noted.

Note 6: NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

Note 7: Optical rotation values for other wavelengths provided in accordance with the formula ICUMSA SPS-1.

Note 8: Where α is the measured angle of optical rotation.

2024-03-06 through 2025-03-31

Effective dates



For the National Voluntary Laboratory Accreditation Program